


STATUTORY DECLARATION Registration of Fittings



I, Greg Strom, Engineering Manager / R&D of Rosemount DP Flowmeters located at 5601 N. 71st street, Boulder Colorado 80301

do solemnly declare that the fittings listed hereunder, which are subject to the Safety Codes Act (check one)

- comply with the requirements of ASME B 31.3 which specifies the dimensions, materials of construction, pressure/temperature ratings and identification marking of the fittings, or are not covered by the provisions of a recognized North American standard and are therefore manufactured to comply with...

I further declare that the manufacture of these fittings is controlled by a quality control program which has been verified by the following authority, Quality Systems Registrar as being suitable for the manufacture of these fittings to the stated standard.

Category F

In support of this application, the following information, calculations and/or test data are attached:

Product drawing 01195-1004, design calculations DC38-001, 002 & 003.

DECLARED before me at Boulder in the state of Colorado this 28th day of March, 2017

(print) Susan L. Mencher (a Commissioner of Oaths or Notary Public)

Notary ID 20064034776

(sign) Susan L. Mencher (a Commissioner of Oaths or Notary Public)

My Commission Expires August 29, 2018 (signature of applicant)

For ABSA Office Use Only:

NOTES: see acceptance letter pp.

To the best of my knowledge and belief, the application meets the requirements of the Safety Codes Act and CSA Standard B51, Clause 4.2, and is accepted for registration in Category F

Registration Number: OF 13185.2 (Signature of the Administrator/SCO)

Date Registered: MAY 04 2017 Expiry Date: Mar. 28, 2023

May 04, 2017

**Attention:** Bill Ringer  
ROSEMOUNT DP FLOW AND OPERATIONS  
5601 N 71 STREET  
BOULDER, CO 80301

**Email:** Bill.Ringer@emerson.com

The design submission, tracking number 2017-02265, originally received on April 13, 2017 was surveyed and accepted for registration as follows:

**CRN :** 0F13185.2 **Accepted on:** May 04, 2017  
**Reg Type:** Addition to Acc. Fitting **Expiry Date:** March 28, 2023  
**Drawing No. :** 21731-001/-004/-005/-006 (SCOPE PER AB-257) Rev 0 As Noted  
**Fitting type:** MODEL 1195 ORIFICE METERS, SIZES 1/2", 1", & 1 1/2"

**The registration is conditional on your compliance with the following notes:**

- The orifice body material is limited to the A351 CF-8M specification only.
- It is our understanding that the flanges used on the 1195 Integrally Orifice Meters shall be in full compliance with the B16.5 Standard.
- The pressure-temperature ratings are as per tables on the drawings.

As indicated on AB-41 Statutory Declaration form and submitted documentation, the code of construction is ASME B31.3.

This registration is valid only for fittings fabricated at the location(s) covered by the QC certificate attached to the accepted AB-41 Statutory Declaration form.


This registration is valid only until the indicated expiry date and only if the Manufacturer maintains a valid quality management system approved by an acceptable third-party agency until that date.

Should the approval of the quality management system lapse before the expiry date indicated above, this registration shall become void.

An invoice covering survey and registration fees will be forwarded from our Revenue Accounts.

Enclosed are stamped prints for your reference.

Sincerely,



RADISAVLJEVIC, ZANA, R.E.T.